

ABSTRACT OF THE DISCLOSURE

A non-contact tonometer changes a reference signal used for determining reliability of a signal indicating deformation of a cornea based on
5 the reflectance of the cornea of an eye to be examined. Thus, the non-contact tonometer can perform stable measurement irrespective of the reflectance of the cornea of the eye to be examined.